



Product Service

EC-Attestation of Compliance

No. EBN 07 09 55375 509

Holder of Certificate:	Electronics Technology Center, 8800 East Foothaven, 1141-L Rosemead, Orange, County of, California-91768-4111-USA	
Name of Object:	Electrical eq. for measurement, control and laboratory use Alcohol tester	
Model(s):	AL-7000	
Description of Object:	Rated input voltage	24 V DC
Tested according to:	EN 61326-1:2007+A1:2008+A2:2011+A3:2013	

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the tested apparatus complies with all essential requirements of the EMC directives and applies only to the sample and its technical documentation submitted to TUV SUD Product Service (center for testing and certification). See also notes overleaf.

Test report no. **CPD00188427**

Date: 2017-09-03



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant procedures have to be observed.

Page 1 of 1